

HART® Field Device Specification
Baumer CombiLyz AFI4/AFI5
Revision 1

Document 81178189, rev. 1

Initial release: 8th of September 2015
Current release: 8th of September 2015

Author: Bo Wellejus Simmons
Baumer A/S
Runetoften 19
DK-8210
DENMARK

® HART is a registered trademark of the HART Communication Foundation

Baumer Document No. 81178189 01
Document Title: CombiLyz AFI4/AFI5 HART Field Device Specification - Revision 1.0
Release Date: 8th of September, 2015

HART FDS
Page 1 of 36

1 Table of Contents

1	Table of Contents	2
2	Introduction	4
2.1	Scope.....	4
2.2	Purpose.....	4
2.3	Who should use this document?.....	4
2.4	Abbreviations and definitions.....	4
2.5	References.....	4
3	Device Identification	4
4	Product Overview	5
5	Product Interfaces	6
5.1	Process Interface.....	6
5.1.1	Sensor Input Channels.....	6
5.2	Host interface.....	6
5.2.1	Analog Output 1: Process Conductivity / Concentration.....	6
5.2.2	Analog Output 2: Process Temperature.....	6
5.3	Local Interfaces.....	6
5.3.1	Local Controls and Displays.....	6
6	Device Variables	6
7	Dynamic Variables	7
8	Status Information	8
8.1	Device Status.....	8
8.2	Extended Device Status.....	8
8.3	Additional Device Status (Command #48).....	9
9	Universal Commands	11
9.1	Supported Universal Commands.....	11
10	Special notes on Universal Commands	11
11	Common-Practice Commands	12
11.1	Supported Common Practice Commands.....	12
11.2	Special notes on Common Practice Commands.....	12
12	Conductivity Family Commands	13
12.1	Supported Conductivity Family Commands.....	13
12.1.1	General Conductivity Device Family Commands (Read).....	13
12.1.2	Programmable Range Conductivity Device Family Commands (Read).....	13
12.1.3	General Conductivity Device Family Commands (Write).....	13
12.1.4	Programmable Range Conductivity Device Family Commands (Write).....	13
12.2	Command #2304 – Read Conductivity Status.....	14
12.3	Command #2305 – Read Capabilities.....	14
12.4	Command #2306 – Read Supported Status Mask.....	14
12.5	Command #2307 – Read Conductivity Sensor Information.....	15
12.6	Command #2308 – Read Device Variable Range and Label Capabilities.....	15
12.7	Command #2309 – Read Conductivity Device Variables.....	16
12.8	Command #2336 – Read Range Hardware Range.....	16
12.9	Command #2337 – Read Range Output Zoom.....	17
12.10	Command #2338 – Read Range Temperature Compensation.....	17
12.11	Command #2339 – Read Range Conductivity Medium Label.....	18
12.12	Command #2340 – Read Range Concentration Table Points.....	19
12.13	Command #2341 – Read Range Concentration Medium Label.....	20
12.14	Command #2342 – Read Range Forced Status.....	20
12.15	Command #2435 – Write Conductivity Sensor Information.....	21
12.16	Command #2464 – Write Range Hardware Range.....	21

12.17	Command #2465 – Write Range Output Zoom	22
12.18	Command #2466 – Write Range Temperature Compensation.....	23
12.19	Command #2467 – Write Range Conductivity Medium Label.....	23
12.20	Command #2468 – Write Range Concentration Table Points	25
12.21	Command #2469 – Write Range Concentration Medium Label.....	26
12.22	Command #2470 – Write Range Forced Status.....	26
13	Device-Specific Commands.....	27
13.1	Command #130 – Read Device Variable State	27
13.2	Command #131 – Write Device Variable State	27
13.3	Command #132 – Read Device Variable Output Channel Setup.....	28
13.4	Command #133 – Write Device Variable Output Channel Setup.....	28
13.5	Command #134 – Reset Configuration to Factory Settings	29
13.6	Command #135 – Trim Device Variable Zero.....	29
13.7	Command #136 – Trim Device Variable Gain	29
13.8	Command #137 – Reset Device Variable Trim.....	29
14	Conductivity Family Device Specific Tables	30
14.1	Cond. Family Device Spec. Table 1 – Conductivity Device Family Device Variable Status	30
14.2	Conductivity Family Device Specific Table 2 – Conductivity Status 0	30
14.3	Conductivity Family Device Specific Table 3 – Conductivity Family Capabilities 0	30
14.4	Conductivity Family Device Specific Table 4 - Conductivity Family Capabilities 1	31
15	Device Specific Tables	31
15.1	Sensor Type Codes	31
15.2	Unit Codes	31
15.3	Internal Unit Conversion	31
16	Performance	31
16.1	Sampling Rates.....	31
16.2	Power-Up	32
16.3	Reset.....	32
16.4	Self-Test.....	32
16.5	Command Response Times	32
16.6	Busy and Delayed-Response	32
16.7	Long Messages.....	32
16.8	Non-Volatile Memory	32
16.9	Modes	32
16.10	Burst Mode.....	32
16.11	Write Protection	32
16.12	Catch Device Variable.....	32
16.13	Damping.....	33
Annex A.	Capability Checklist.....	34
Annex B.	Default Configuration	35
Annex C.	Revision History	36

2 Introduction

2.1 Scope

Baumer CombiLyz AFI4/AFI5 HART conductivity transmitter has built-in support for the HART 7.4 protocol. Since the HART version of the transmitter operates identically with the non-HART version of the transmitter, this document focuses solely on the HART functionalities of the transmitter. For all other operational aspects of the transmitter, please consult the data sheet and the user guide.

This document contains the necessary data for an operator, familiar with the HART protocol, to access all functions of the transmitter from a master system.

2.2 Purpose

This specification is designed to complement other documentation (e.g., the *AFI4/AFI5 User Guide*) by providing a complete, unambiguous description of this Field Device from a HART Communication perspective

2.3 Who should use this document?

The specification is designed to be a technical reference for HART capable Host Application Developers, System Integrators and knowledgeable End Users. It also provides functional specifications (e.g., commands, enumerations and performance requirements) used during Field Device development, maintenance and testing. This document assumes the reader is familiar with HART Protocol requirements and terminology.

2.4 Abbreviations and definitions

CT	Common Table
DT	Device Specific Table
uint-8	8-bit unsigned integer, representing value 0 .. 255, can also be used for single bit flags
uint-16	16-bit unsigned integer, representing value 0 .. 65535
float	32-bit IEEE-754 (IEC 559) compatible single floating point variable
ASCII	ISO Latin-1 (ISO 8859) string text
packed	HART specific Packed ASCII format
PV	Primary Variable
SV	Secondary Variable
TV	Tertiary Variable
QV	Quaternary Variable

2.5 References

HART Smart Communications Protocol Specification. HCF_SPEC-12. Available from the HCF.
CombiLyz AFI4/AFI5, Operating Instruction, Document 11126179 05. Available from www.baumer.com.

3 Device Identification

Manufacturer Name:	Baumer	Model Name(s):	AFI4/AFI5 CombiLyz
Manufacture ID Code:	96 (60 Hex)	Device Type Code:	236 (EC Hex)
HART Protocol Revision	7.4	Device Revision:	1
Number of Device Variables	8		
Physical Layers Supported	FSK		
Physical Device Category	Current Output		

4 Product Overview

The CombiLyz AFI4/AFI5 is an externally powered 4-20mA current output conductivity transmitter, with an extra current output channel for temperature. The HART access to transmitter is via the conductivity current output channel.

The conductivity current output can be configured to output a concentration value for up to 4 different configurable media, internally calculated from the measured conductivity value.

Temperature compensation is fully configurable, and the temperature used for this is either the internally measured sensor tip temperature, fixed or provided by the HART master.

A 2-pin discrete input channel allows selecting 1 of 4 programmable ranges, which allows runtime switching between e.g. different zoom ranges or different concentration media. Selecting range with a HART command is also possible, overriding discrete input on these pins.

5 Product Interfaces

5.1 Process Interface

5.1.1 Sensor Input Channels

The main conductivity sensor is mounted around the physical hole in the sensor.

A temperature sensor is located in the very tip of the sensor, providing an accurate temperature reading of the media temperature, which can be used for temperature compensation, as well as being output on the temperature current output channel.

5.2 Host interface

The transmitter has two 4-20mA current output channels. Each of these can be independently configured with linear over-range from 3.5 to 23 mA. The Conductivity current output channel supports HART Communication. The error output current, to be used in case of transmitter error, is fully configurable in the range 3.5 to 23 mA, as well as which channel (or both) the error current is to be output on. The error current value is common for both channels, and cannot overlap the operating output current ranges of the selected error output channels.

	Direction	Values (% of range)	Values (mA or V)
Linear over-range	Down	-3.13% to -0.01%	3.50 to 3.99 mA
	Up	100.01% to 118.75%	20.01 to 23.00 mA
Device malfunction indication	Fixed value	-3.13% to 118.75%	3.50 to 23.00 mA
Maximum current		118.75%	23.0 mA
Multi-Drop current draw			4.0 mA
Lift-off voltage			15.0 V

5.2.1 Analog Output 1: Process Conductivity / Concentration

The two-wire 4-20mA conductivity loop current output is connected on two terminals marked "4-20mA/Cond/Conc" (+/-). Refer to the Operating Instruction for connection details.

This output corresponds to Device Variable 4, which is the device variable for combined Conductivity and Concentration, depending on which operating mode is selected.

HART Communication is supported on this current loop output.

5.2.2 Analog Output 2: Process Temperature

The two-wire 4-20mA temperature loop current output is connected on two terminals marked "4-20mA/Temp" (+/-). Refer to the Operating Instruction for connection details.

This output corresponds to Device Variable 1, which is the device variable for the process temperature.

HART Communication is NOT supported on this current loop output.

5.3 Local Interfaces

5.3.1 Local Controls and Displays

This device can be attached to a DFON display, providing local in-situ measurement readout and configuration possibilities. A DFON display is connected with the provided flat ribbon cable provided with the display unit. FlexProgram configuration is also possible on the two Com. terminals. A PC and a Baumer FlexProgrammer 9701 must be used for this. Standalone configuration with the FlexProgrammer is not possible on the AFI4/AFI5.

A 2-pin discrete input channel exists for selecting configurable ranges.

Please refer to the Operating Instruction manual for more information.

6 Device Variables

This Field Device does not expose any Device Variables.

DV No.	Name	Description	Unit codes	Classification code
0, 246	Conductivity	Conductivity measurement with	66 mS/cm	87 Conductance

	PV	temperature compensation Current Output capability on current output channels: 2 (indirectly via Dev. Var. 4)	67	$\mu\text{S/cm}$	
1, 247	Temperature SV	Sensor tip temperature measurement Current Output capability on current output channels: 1	32 33	$^{\circ}\text{C}$ $^{\circ}\text{F}$	64 Temperature
2, 248	Concentration TV	Concentration value calculated from the Conductivity measurement Current Output capability on current output channels: 2 (indirectly via Dev. Var. 4)	57	%	90 Concentration
3, 249	Raw Conductivity QV	Conductivity measurement without temperature compensation	66	mS/cm	87 Conductance
4	Concentration / Conductivity	Mode dependent. Value, unit and classification fetched directly from either Device Variable 0 or Device Variable 2. Current Output capability on current output channels: 2	66 67 57	mS/cm $\mu\text{S/cm}$ %	87 Conductance 90 Concentration
6, 244	Percent of Range	Output in % of full scale	57	%	0 Not Class'd
7, 245	Loop Current	Loop Current associated with Device Variable 4, representing either Conductivity or Concentration	39	mA	84 Current
8	Loop Current (Secondary)	Loop Current associated with Device Variable 1, representing sensor temperature	39	mA	84 Current

Only Conductivity (PV) and Temperature (SV) allow changing of unit codes. Device Variable 4 automatically adapts the used Conductivity unit code.

7 Dynamic Variables

Two Dynamic Variables are implemented.

Dyn var.	Meaning	Units
PV	Conductivity	$\mu\text{S/cm}$, mS/cm
SV	Temperature	$^{\circ}\text{C}$, $^{\circ}\text{F}$
TV	Concentration	%
QV	Raw Conductivity	mS

8 Status Information

8.1 Device Status

The Field Device Status byte is contained in the second data byte in messages from the device. The following table defines the meaning of the different status bits.

Bit	Definition	Description
7	Device Malfunction	Is set if an electronic defect or memory defect is detected.
6	Configuration Changed	Is set if a HART command results in writing new data to a configuration register. Writing a new value identical to the currently stored value, does not result in this bit being set.
5	Cold Start	Is set upon restart. It is reset for each master after responding to the first command from that specific master.
4	More Status Available	Is set if any of the Additional Device Status bits change status.
3	Loop Current Fixed	This bit is set if device is running in Fixed Current Mode (Command 40) or if Loop Current Signaling mode is turned off (e.g. in Multidrop Mode).
2	Loop Current Saturated	Is set if the output current associated to PV/TV/DV4 is capped by either the upper or lower current limits.
1	Non-Primary Variable Out of Limits	Is set if either SV, TV or QV is high or low limited.
0	Primary Variable Out of Limits	Is set if PV is high or low limited.

8.2 Extended Device Status

Extended Device Status is returned along with Additional Device Status by HART Command 48. Two bits are supported in this device.

Bit	Definition	Description
2	Critical Power Failure	Is set if the device detects that the power supply is not performing as expected.
1	Device Variable Alert	This bit is set if any Device Variable is simulated/fixed, or the environmental conditions are out of range. It will also be set if an electronic defect or memory defect is detected.

8.3 Additional Device Status (Command #48)

Command #48 returns 14 bytes of data, with the following status information:

Byte	Bit	Definition	Description
0	Device Specific Error Status Flags		
	7-1	NOT USED	
	0	General Transmitter Error	Is set if wire break is detected
1	Device Specific 0		
	7-0	NOT USED	
2	Device Specific 1		
	7-0	NOT USED	
3	Device Specific 2		
	7-0	NOT USED	
4	Device Specific 3		
	7-0	NOT USED	
5	Device Specific 4		
	7-0	NOT USED	
6	Extended Device Status		
	7-3	NOT USED	
	2	Critical Power Failure	See Extended Device Status
	1	Device Variable Alert	See Extended Device Status
	0	NOT USED	
7	NOT USED		
	7-0	NOT USED	
8	Standardized Status 0		
	7	NOT USED	
	6	Electronic Defect	Is set in case of wire break
	5	Environment Conditions out of Range	This bit is set if either media or ambient temperatures out of range.
	4	Power Supply Conditions out of Range	Is set if the device detects that the power supply is not performing as expected.
	3	Watchdog Reset Executed	This bit is set in case of the watchdog resetting the device, in case of firmware running into a software dead-lock.
	2	Non-volatile Memory Defect	This bit is set if a problem with the system memory is detected.
	1	Device Variable Simulation Active	Is set if any device variable is being simulated, e.g. by in-factory system test.
	0	NOT USED	
9	Standardized Status 1		
	7-3	NOT USED	
	2	Event Notification Overflow	This bit is set if the internal processor becomes overworked, not able to execute all tasks given within the allowed time.
	1	Discrete Variable Simulation	Is set if any device variable is being simulated, e.g. by in-factory system test.
	0	Simulation Active	Is set if any device variable is being simulated, e.g. by in-factory system test.

Table continues on next page ..

.. table continued from previous page.

10	Analog Channel Saturated		
	7-3	NOT USED	
	2	Analog Channel 2	Is set if Analog Channel 2 is capped by either the upper or lower current limit.
	1	Analog Channel 1	Is set if Analog Channel 1 is capped by either the upper or lower current limit.
	0	NOT USED	
11	Standardized Status 2		
	7-0	NOT USED	
12	Standardized Status 3		
	7-0	NOT USED	
13	Analog Channel Fixed		
	7-3	NOT USED	
	2	Analog Channel 2	Is set if Analog Channel 2 is fixed by in-factory system test.
	1	Analog Channel 1	Is set if Analog Channel 1 is fixed by either Fixed Current Mode (Command 40) or if Loop Current Signaling mode is turned off (e.g. in Multidrop Mode). It can be caused by a running in-factory system test.
	0	NOT USED	

NOT USED bits are always set to 0.

These status bits are updated several times each second. They are set by any failure detected by the periodic status update routine.

9 Universal Commands

9.1 Supported Universal Commands

All Universal Commands are mandatory and are supported by the device. Following Universal Commands are implemented:

0	Read Unique Identifier
1	Read Primary Variable
2	Read Loop Current And Percent Of Range
3	Read Dynamic Variables And Loop Current
6	Write Polling Address
7	Read Loop Configuration
8	Read Dynamic Variable Classifications
9	Read Device Variables with Status
11	Read Unique Identifier Associated With Tag
12	Read Message
13	Read Tag, Descriptor, Date
14	Read Primary Variable Transducer Information
15	Read Device Information
16	Read Final Assembly Number
17	Write Message
18	Write Tag, Descriptor, Date
19	Write Final Assembly Number
20	Read Long Tag
21	Read Unique Identifier Associated With Long Tag
22	Write Long Tag
38	Reset Configuration Changed Flag
48	Read Additional Device Status

10 Special notes on Universal Commands

Command #3: Returns PV, SV, TV and QV. This totals in 24 data bytes.

Command #9: This command supports up to 4 device variables. This totals in up to 37 data bytes, including the time stamp.

If more than 4 device variables are requested, only the first 4 are returned, along with a warning.

Command #14: Transducer serial number is not supported. The units code for limits and minimum span is equal to that of the Primary Variable.

Since the physical transducer can be switched between different physical ranges, the returned transducer limits reflect those of the currently selected range's limits.

Command #15: Write protect is not implemented, and Write Protect Code is therefore always returned as "251" (None). The unit code for Primary Variable range values is the same as is used for the Primary Variable.

11 Common-Practice Commands

11.1 Supported Common Practice Commands

The following common-practice commands are implemented:

34	Write Primary Variable Damping Value
35	Write Primary Variable Range Values
36	Set Primary Variable Upper Range Value
37	Set Primary Variable Lower Range Value
38	Reset "Configuration Changed" Flag
40	Enter/Exit Fixed Current Mode
42	Perform Device Reset
43	Set Primary Variable Zero
44	Write Primary Variable Units
45	Trim Loop Current Zero
46	Trim Loop Current Gain
48	Read Additional Device Status
50	Read Dynamic Variable Assignment
53	Write Device Variable Units

11.2 Special notes on Common Practice Commands

Command #35: Please see *Common note 1*.

Command #36: Please see *Common note 1*.

Command #37: Please see *Common note 1*.

Command #45: Prior to issuing this command, the loop current must be fixed at exactly 4.000mA (set with command 40).

Command #46: Prior to issuing this command, the loop current must be fixed at exactly 20.000mA (set with command 40).

Command #48: Returns 14 bytes of data.

Common note 1: Since this transmitter has several different physical measurement ranges, these commands only allows execution when the first range (Range 1) is selected either by discrete input pins R1 and R2 or by the designated Conductivity Family HART command (Command 2470). If any other range is being used, this command will respond with error code 15 – “Non-Active Range Attempted Changed”.

Additionally, this device does not support reverse operation. This means that PV Upper Range Value must be bigger than PV Lower Range Value.

12 Conductivity Family Commands

The following conductivity family commands are taken from the new preliminary Conductivity Family Specification from the active Conductivity Working Group, since the original preliminary Conductivity Family Specification contains flawed command numbers. Additionally the original preliminary lacked vital features regarding ranges, used by many conductivity transmitter types.

Command description for all these new commands is added here, to minimize confusion of command specification is which.

The upper byte of all the 16-bit command codes is set to 0x09 for Conductivity Family Commands, and in theory there are command codes enough for 256 commands in the family.

For the sake of simplicity, all read commands are located in the range 0x0900 to 0x097F, while write commands are located from 0x0980 to 0x09FF.

Additionally, the commands are divided the following 2 groups:

General Conductivity Device Family Commands (Read and Write)

These commands are much like commands for other device families, reading device information and status, as well as giving the master an overview of the capabilities of the device.

Programmable Range Conductivity Device Family Commands (Read and Write)

Many conductivity devices allow selecting different ranges at runtime, either by a discrete input, or via HART commands. These ranges most often can be programmed differently, ex. to perform special hardware zoom in certain situations, or ex. to select between concentration media setup. This section contains commands to read and write the setup of variables specifically concerning these ranges.

12.1 Supported Conductivity Family Commands

Implemented conductivity family commands are as follows:

12.1.1 General Conductivity Device Family Commands (Read)

2304	Read Conductivity Status
2305	Read Capabilities
2306	Read Supported Status Mask
2307	Read Conductivity Sensor Information
2308	Read Device Variable Range and Label Capabilities
2309	Read Conductivity Device Variables

12.1.2 Programmable Range Conductivity Device Family Commands (Read)

2336	Read Range Hardware Range
2337	Read Range Output Zoom
2338	Read Range Temperature Compensation
2339	Read Range Conductivity Medium Label
2340	Read Range Concentration Table Points
2341	Read Range Concentration Medium Label
2342	Read Range Forced Status

12.1.3 General Conductivity Device Family Commands (Write)

2435	Write Conductivity Sensor Information
------	---------------------------------------

12.1.4 Programmable Range Conductivity Device Family Commands (Write)

2464	Write Range Hardware Range
2465	Write Range Output Zoom
2466	Write Range Temperature Compensation
2467	Write Range Conductivity Medium Label
2468	Write Range Concentration Table Points
2469	Write Range Concentration Medium Label
2470	Write Range Forced Status

12.2 Command #2304 – Read Conductivity Status

This command reads out status for the specified Pressure Device Variable

Request data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)

Response data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)
1	uint-8	Conductivity Family Device Variable Status (FT 1)
2	uint-8	Conductivity Family Status 0 (FT 2)

Command specific response codes for command 2304

Code	Type	Description
17	Error	Invalid Device Variable Index
19	Error	Device Variable Index Not Allowed For This Command

12.3 Command #2305 – Read Capabilities

This command reads the capabilities of the Conductivity Device Family Variable. Some functions are optional and this command shows which ones are implemented. Each bit in the capability registers describes a feature.

Request data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)

Response data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)
1	uint-8	Family Definition Revision
2	uint-8	Conductivity Family Capabilities 0 (FT 3)
3	uint-8	Conductivity Family Capabilities 1 (FT 4)

Command specific response codes for command 2305

Code	Type	Description
17	Error	Invalid Device Variable Index
19	Error	Device Variable Index Not Allowed For This Command

12.4 Command #2306 – Read Supported Status Mask

This command reads out a bit mask, used to identify which corresponding status flags (read using Command #2304 – Read Conductivity Status) are supported by the device.

Request data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)

Response data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)
1	uint-8	Supported Device Family Variable Status Bit Mask (FT 1)
2	uint-8	Supported Conductivity Status 0 Bit Mask (FT 2)

Command specific return codes for command 2306

Code	Type	Description
17	Error	Invalid Device Variable Index
19	Error	Device Variable Index Not Allowed For This Command

12.5 Command #2307 – Read Conductivity Sensor Information

This command reads out the sensor information of the physical transducer used to get data for use with the specified device variable.

Request data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)

Response data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)
1	uint-8	Sensor Type
2..5	float	Cell Constant

Command specific response codes for command 2307

Code	Type	Description
17	Error	Invalid Device Variable Index
19	Error	Device Variable Index Not Allowed For This Command

Cell constant is not the accurate cell constant, but the typical cell constant of this cell type. The cell constant is given in the unit (cm⁻¹).

12.6 Command #2308 – Read Device Variable Range and Label Capabilities

This command reads the Range capabilities for the specified device variable – which give an impression of how to configure ranges and labels.

Number of ranges specify how many ranges this device can switch between.

Number of labels specify how many labels are available in total for this device variable.

Number of labels for each range specify how many label are unique for each range for this device variable.

Range flags define some more special details for the device variables.

Request data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)

Response data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)
1	uint-8	Number of ranges
2	uint-8	Number of labels associated with this device variable
3	uint-8	Number of labels associated with each Range for this Device Variable
4	uint-8	Range flags (DT 10)

Command specific response codes for command 2308

Code	Type	Description
17	Error	Invalid Device Variable Index
19	Error	Device Variable Index Not Allowed For This Command

12.7 Command #2309 – Read Conductivity Device Variables

This command reads the all the standard conductivity device variables Conductivity Device Family Variable. This standard conductivity device command for some reason uses a device variable code to output standard values.

Request data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)

Response data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)
1..4	float	Conductivity
5..8	float	Temperature
9..12	float	Concentration
13..16	float	Raw Conductivity

Command specific response codes for command 2309

Code	Type	Description
17	Error	Invalid Device Variable Index
19	Error	Device Variable Index Not Allowed For This Command

CombiLyz AFI4/AFI5 specific:

Only Device Variable Code 0 (and Primary Variable Code 245) is allowed in this device! The function therefore always returns Device Variable 0, 1, 2 and 3, which are Conductivity, Temperature, Concentration and Raw Conductivity.

12.8 Command #2336 – Read Range Hardware Range

This command reads the hardware range of the specified range.

Request data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)
1	uint-8	Range number (DT 2)

Response data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)
1	uint-8	Range number (DT 2)
2	uint-8	Range unit (CT 2)
3..6	float	Lower hardware range input (DT 3)
7..10	float	Upper hardware range input (DT 4)

Command specific response codes for command 2336

Code	Type	Description
15	Error	Invalid Programmable Range Number
17	Error	Invalid Device Variable Index
19	Error	Device Variable Index Not Allowed For This Command

Range unit defines the unit in which the values are returned.

CombiLyz AFI4/AFI5 specific:

Only usable on the Conductivity Family Device Variables. All device variables share the same ranges.

12.9 Command #2337 – Read Range Output Zoom

This command reads the zoom of the selected range.

The output is adjusted according to the following scaling: Lower range zoom value correspond to the Lower range loop output current value, and Upper range zoom value corresponding to the Upper range loop current value.

Request data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)
1	uint-8	Range number (DT 2)

Response data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)
1	uint-8	Range number (DT 2)
2	uint-8	Range unit (CT 2)
3..6	float	Lower range zoom
7..10	float	Upper range zoom
11..14	float	Lower range loop output current [mA]
15..18	float	Upper range loop output current [mA]

Command specific response codes for command 2337

Code	Type	Description
15	Error	Invalid Programmable Range Number
17	Error	Invalid Device Variable Index
19	Error	Device Variable Index Not Allowed For This Command

Range unit specifies the unit in which the values are returned.

12.10 Command #2338 – Read Range Temperature Compensation

This command reads the temperature compensation variables for the selected range.

Request data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)
1	uint-8	Range number (DT 2)

Response data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)
1	uint-8	Range number (DT 2)
2	uint-8	Temperature unit (CT 2)
3	uint-8	Temperature source (DT 7)
4..7	float	Compensation temperature
8..11	float	Linear compensation [%/K]
12..15	float	Exponential compensation [%/K ²]
16..19	float	Reference temperature

Command specific response codes for command 2338

Code	Type	Description
15	Error	Invalid Programmable Range Number
17	Error	Invalid Device Variable Index
19	Error	Device Variable Index Not Allowed For This Command

Compensation temperature returned is the temperature used, from the specified source.

The Temperature unit is not necessarily the temperature unit used in the device, but the unit for the temperature values returned with this command.

12.11 Command #2339 – Read Range Conductivity Medium Label

Many devices are able to assume the presence of a specific medium, if the conductivity is within a certain range.

This command reads the label string which is configured for the specified conductivity range.

Use conductivity label no. 0 (zero) to read the label used when outside of the other ranges.

Request data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)
1	uint-8	Range number (DT 2)
2	uint-8	Conductivity label no. (DT 5)

Response data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)
1	uint-8	Range number (DT 2)
2	uint-8	Conductivity label no. (DT 5)
3	uint-8	Conductivity unit (CT 2)
4..7	float	Lower limit for conductivity media label range
8..11	float	Upper limit for conductivity media label range
12..27	ASCII	Conductivity media label

Command specific response codes for command 2339

Code	Type	Description
13	Error	Invalid Label Number
15	Error	Invalid Programmable Range Number
17	Error	Invalid Device Variable Index
19	Error	Device Variable Index Not Allowed For This Command

NOTE:

This command can only be used on Device Variables with Conductivity Classification.

CombiLyz AFI4/AFI5 specific:

This command can only be used on the Primary Device Variable.

All Ranges share the same media labels.

12.12 Command #2340 – Read Range Concentration Table Points

This command reads the specified Conductivity/Concentration table points. It also returns the number of table points to use in total for this range. It is possible to read up to 8 table points with each issue of the command. Point number specifies the first point to be read with this command, while Number of table points in message defines how many points are to be read with the command.

It is optional for a device to support more than one point for each command. It must be clearly stated in the device HART documentation how many points are supported for command 2340, or if a very specific number of points are to be used.

Note: Command 2340 and Command 2468 must support the same amount of points per message.

Request data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)
1	uint-8	Range number (DT 2)
2	uint-8	Point number (point x)
3	uint-8	Number of table points in msg.

Response data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)
1	uint-8	Range number (DT 2)
2	uint-8	Point number (point x)
3	uint-8	Number of table points in msg.
4	uint-8	Number of table points to use
5	uint-8	Conductivity unit
6	uint-8	Concentration unit
7..10	float	Point x Conductivity value
11..14	float	Point x Concentration value
15..18	float	Point x+1 Conductivity value
19..22	float	Point x+1 Concentration value
23..26	float	Point x+2 Conductivity value
27..30	float	Point x+2 Concentration value
31..34	float	Point x+3 Conductivity value
35..38	float	Point x+3 Concentration value
39..42	float	Point x+4 Conductivity value
43..46	float	Point x+4 Concentration value
47..50	float	Point x+5 Conductivity value
51..54	float	Point x+5 Concentration value
55..58	float	Point x+6 Conductivity value
59..62	float	Point x+6 Concentration value
63..66	float	Point x+7 Conductivity value
67..70	float	Point x+7 Concentration value

Command specific response codes for command 2340

Code	Type	Description
2	Error	Invalid Selection
15	Error	Invalid Programmable Range Number
17	Error	Invalid Device Variable Index
19	Error	Device Variable Index Not Allowed For This Command
29	Error	Invalid Table Point Selection

NOTE: The parameters with the grey background fill are only returned if at least one table point is read.

This command can only be used on Device Variables with Concentration Classification.

CombiLyz AFI4/AFI5 specific:

Points are denoted 1 .. 30.

Limited to 4 points in each message.

12.13 Command #2341 – Read Range Concentration Medium Label

This command reads the concentration setup. It reads concentration mode status for the specified Concentration Device Variable, and reads the label string for the specified Range number.

Request data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)
1	uint-8	Range number (DT 2)

Response data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)
1	uint-8	Range number (DT 2)
2..17	ASCII	Concentration range label

Command specific response codes for command 2341

Code	Type	Description
15	Error	Invalid Programmable Range Number
17	Error	Invalid Device Variable Index
19	Error	Device Variable Index Not Allowed For This Command

This command can only be used on Device Variables with Concentration Classification.

12.14 Command #2342 – Read Range Forced Status

The active range is often selectable using a combination of + and - on digital input pins. This command reads the status of the forced range used. NOT_USED is returned if it not forced. This setting is volatile, and will be set to NOT_USED on reset!

Request data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)

Response data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)
1	uint-8	Range number (DT 2)

Command specific response codes for command 2342

Code	Type	Description
17	Error	Invalid Device Variable Index
19	Error	Device Variable Index Not Allowed For This Command

CombiLyz AFI4/AFI5 specific:

This command reads the forced range for all Device Variables, no matter which Device Variable is targeted. Only one range is selectable at a time for all Device Variables.

12.15 Command #2435 – Write Conductivity Sensor Information

This command writes sensor type and cell constant.

Request data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)
1	uint-8	Sensor Type
2..5	float	Cell Constant

Response data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)
1	uint-8	Sensor Type
2..5	float	Cell Constant

Command specific response codes for command 2435

Code	Type	Description
2	Error	Invalid Selection (Sensor Type) (Not relevant)
3	Error	Passed Parameter Too Large (Cell Constant) (Not relevant)
4	Error	Passed Parameter Too Small (Cell Constant) (Not relevant)
16	Error	Access Restricted
17	Error	Invalid Device Variable Index
19	Error	Device Variable Index Not Allowed For This Command

CombiLyz AFI4/AFI5 specific:

Changing sensor type and cell constant is not allowed on this device. Consequently, all correct commands always return return-code ACCESS_RESTRICTED!

Cell constant is not the accurate cell constant, but the typical cell constant of this cell type. The cell constant is given in the unit (cm^{-1}).

12.16 Command #2464 – Write Range Hardware Range

This command writes the hardware range of the selected conductivity range.

Request data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)
1	uint-8	Range number (DT 2)
2	uint-8	Range unit (CT 2)
3..6	float	Lower hardware range input (DT 3)
7..10	float	Upper hardware range input (DT 4)

Response data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)
1	uint-8	Range number (DT 2)
2	uint-8	Range unit (CT 2)
3..6	float	Lower hardware range input (DT 3)
7..10	float	Upper hardware range input (DT 4)

Command specific response codes for command 2464

Code	Type	Description
8	Warning	Set To Nearest Value
15	Error	Invalid Programmable Range Number
17	Error	Invalid Device Variable Index
18	Error	Invalid Units Code
19	Error	Device Variable Index Not Allowed For This Command

Range unit defines the unit in which the values are sent and returned. It does not setup the unit for the device variable.

To make it easier to find a valid hardware range, then the device should automatically set the range to the closest valid values, preferably with a bigger span, not truncating the requested hardware range.

CombiLyz AFI4/AFI5 specific:

Modifying Lower conductivity range input and Upper range input for the Conductivity variable zoom range will be limited by the hardware range selected with this command.

Modifying the hardware range will also set the Output Zoom range to this specific value.

This can be manually set afterwards using command Command #2465 – Write Range Output Zoom.

If invalid value for either Lower Hardware Range Input or Lower Hardware Range Input is invalid, it will be set to nearest value, see Device Specific Table 3 and 4.

12.17 Command #2465 – Write Range Output Zoom

This command writes the output zoom of the selected range for the specified device variable.

The output is adjusted according to the following scaling: Lower range zoom value correspond to the Lower range loop output current value, and Upper range zoom value corresponding to the Upper range loop current value.

The units allowed depend on the device variable number (and type).

Request data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)
1	uint-8	Range number (DT 2)
2	uint-8	Range unit (CT 2)
3..6	float	Lower range zoom
7..10	float	Upper range zoom
11..14	float	Lower range loop output current [mA]
15..18	float	Upper range loop output current [mA]

Response data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)
1	uint-8	Range number (DT 2)
2	uint-8	Range unit (CT 2)
3..6	float	Lower range zoom
7..10	float	Upper range zoom
11..14	float	Lower range loop output current [mA]
15..18	float	Upper range loop output current [mA]

Command specific response codes for command 2465

Code	Type	Description
9	Error	Lower Range Value Too High
10	Error	Lower Range Value Too Low
11	Error	Upper Range Value Too High
12	Error	Upper Range Value Too Low
13	Error	Invalid Current Value
15	Error	Invalid Programmable Range Number
17	Error	Invalid Device Variable Index
18	Error	Invalid Units Code
19	Error	Device Variable Index Not Allowed For This Command
29	Error	Invalid Span

If there are no different ranges available (one single large range) for the specified Device Variable, then Range number 0 must be used.

Range unit specifies the unit in which the values are sent. It does not setup the unit for the device variable.

CombiLyz AFI4/AFI5 specific:

Lower range input and Upper range input for the Conductivity range will be limited by the selected hardware range.

Lower range loop output current must always be 4 mA.

Upper range loop output current must always be 20 mA.

If using Command #2340 – Read Range Concentration Table Points then the zoom of the specified Concentration device variable may be set according to the table.

12.18 Command #2466 – Write Range Temperature Compensation

This command writes the temperature compensation variables for the selected range.

Request data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)
1	uint-8	Range number (DT 2)
2	uint-8	Temperature unit (CT 2)
3	uint-8	Temperature source (DT 7)
4..7	float	Compensation temperature
8..11	float	Linear compensation [%/K]
12..15	float	Exponential compensation [%/K ²]
16..19	float	Reference temperature

Response data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)
1	uint-8	Range number (DT 2)
2	uint-8	Temperature unit (CT 2)
3	uint-8	Temperature source (DT 7)
4..7	float	Compensation temperature
8..11	float	Linear compensation [%/K]
12..15	float	Exponential compensation [%/K ²]
16..19	float	Reference temperature

Command specific response codes for command 2466

Code	Type	Description
2	Error	Invalid Selection
3	Error	Passed Parameter Too Large
4	Error	Passed Parameter Too Small
15	Error	Invalid Programmable Range Number
17	Error	Invalid Device Variable Index
18	Error	Invalid Units Code
19	Error	Device Variable Index Not Allowed For This Command
29	Error	Invalid Coefficient

If Exponential compensation is not used, it should be set to zero.

If Reference temperature is not used, it should be set to 25°C, which is standard reference temperature.

The *Compensation temperature* sent to transmitter is ignored if it is set to use internal or HART temperature source. However, if it runs in Fixed temperature mode, the this is used to set the Fixed temperature.

Compensation temperature returned is the temperature used, from the specified source.

The Temperature unit is not the temperature unit to be used in the device, but the unit for the temperature values sent with this command.

CombiLyz AFI4/AFI5 specific:

The only temperature unit allowed for this command is °C

There is no limit for either the temperature and compensation values that can be used for this command, error code Invalid Coefficient is therefore not used on AFIx

Using this function sets the Advanced compensation variable to enabled!

12.19 Command #2467 – Write Range Conductivity Medium Label

Many devices are able to assume the presence of a specific medium, if the conductivity is within a certain range.

This command writes the label string for the specified conductivity range.

Use conductivity label no. 0 (zero) to write the label used when outside of the other ranges.

Request data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)
1	uint-8	Range number (DT 2)
2	uint-8	Conductivity label no. (DT 5)

Response data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)
1	uint-8	Range number (DT 2)
2	uint-8	Conductivity label no. (DT 5)

3	uint-8	Conductivity unit (CT 2)
4..7	float	Lower limit for conductivity media label range
8..11	float	Upper limit for conductivity media label range
12..27	ASCII	Conductivity media label

3	uint-8	Conductivity unit (CT 2)
4..7	float	Lower limit for conductivity media label range
8..11	float	Upper limit for conductivity media label range
12..27	ASCII	Conductivity media label

Command specific response codes for command 2467

Code	Type	Description
9	Error	Lower Range Value Too High
10	Error	Lower Range Value Too Low
11	Error	Upper Range Value Too High
12	Error	Upper Range Value Too Low
13	Error	Invalid Label Number
15	Error	Invalid Programmable Range Number
17	Error	Invalid Device Variable Index
18	Error	Invalid Units Code
19	Error	Device Variable Index Not Allowed For This Command

NOTE:

This command can only be used on Device Variables with Conductivity Classification.

CombiLyz AFI4/AFI5 specific:

This command can only be used on the Primary Device Variable.

This command does the same to all "Ranges", no matter which range is selected.

This command will report failure if writing a label in a range that overlaps another label range. Reduce the size of the range that is in the way to set the new label range.

12.20 Command #2468 – Write Range Concentration Table Points

This command sets up the specified Conductivity/Concentration table points. It also specifies the number of table points to use in total for this range. It is possible to setup up to 8 table points with each issue of the command.

Point number specifies the first point to be changed with this command, while Number of table points in message defines how many points are included in the command message.

It is optional for a device to support more than one point for each command.

If command is issued with 0 points in the message, it is not necessary to send units.

Note: Command 2340 and Command 2468 must support the same amount of points per message.

Request data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)
1	uint-8	Range number (DT 2)
2	uint-8	Point number (point x)
3	uint-8	Number of table points in msg.
4	uint-8	Number of table points to use
5	uint-8	Conductivity unit
6	uint-8	Concentration unit
7..10	float	Point x Conductivity value
11..14	float	Point x Concentration value
15..18	float	Point x+1 Conductivity value
19..22	float	Point x+1 Concentration value
23..26	float	Point x+2 Conductivity value
27..30	float	Point x+2 Concentration value
31..34	float	Point x+3 Conductivity value
35..38	float	Point x+3 Concentration value
39..42	float	Point x+4 Conductivity value
43..46	float	Point x+4 Concentration value
47..50	float	Point x+5 Conductivity value
51..54	float	Point x+5 Concentration value
55..58	float	Point x+6 Conductivity value
59..62	float	Point x+6 Concentration value
63..66	float	Point x+7 Conductivity value
67..70	float	Point x+7 Concentration value

Response data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)
1	uint-8	Range number (DT 2)
2	uint-8	Point number (point x)
3	uint-8	Number of table points in msg.
4	uint-8	Number of table points to use
5	uint-8	Conductivity unit
6	uint-8	Concentration unit
7..10	float	Point x Conductivity value
11..14	float	Point x Concentration value
15..18	float	Point x+1 Conductivity value
19..22	float	Point x+1 Concentration value
23..26	float	Point x+2 Conductivity value
27..30	float	Point x+2 Concentration value
31..34	float	Point x+3 Conductivity value
35..38	float	Point x+3 Concentration value
39..42	float	Point x+4 Conductivity value
43..46	float	Point x+4 Concentration value
47..50	float	Point x+5 Conductivity value
51..54	float	Point x+5 Concentration value
55..58	float	Point x+6 Conductivity value
59..62	float	Point x+6 Concentration value
63..66	float	Point x+7 Conductivity value
67..70	float	Point x+7 Concentration value

Command specific response codes for command 2468

Code	Type	Description
2	Error	Invalid Selection
3	Error	Passed Parameter Too Large (Not implemented)
4	Error	Passed Parameter Too Small (Not implemented)
9	Error	Lower Range Value Too High (Not implemented)
10	Error	Lower Range Value Too Low (Not implemented)
11	Error	Upper Range Value Too High (Not implemented)
12	Error	Upper Range Value Too Low (Not implemented)
15	Error	Invalid Programmable Range Number
17	Error	Invalid Device Variable Index
18	Error	Invalid Units Code
19	Error	Device Variable Index Not Allowed For This Command
28	Error	Command Data Overflow
29	Error	Invalid Table Point Selection

This command can only be used on Device Variables with Concentration Classification.

NOTE: The parameters with the blue background fill are only returned, if at least one table point is written.

CombiLyz AFI4/AFI5 specific:

Points are denoted 1 .. 30.

Limited to 4 points in each message.

If setting Concentration mode Enable for Device Variable 2, Device Variable 4 will output Concentration instead of Conductivity, and the Loop Current output will output the Conductivity output current.

There is no check on the conductivity and concentration values! It can cause erratic behavior if variables are not properly setup. Irrelevant response codes are therefore have a strike-through in the response code list.

12.21 Command #2469 – Write Range Concentration Medium Label

This command writes the label string for the specified Range number.

Please note that this command can be issued with a one byte predefined media label number instead of a full media label string of 16 bytes. In either way, the response always returns the full written data string.

Request data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)
1	uint-8	Range number (DT 2)
2..17	ASCII	Concentration range label

Response data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)
1	uint-8	Range number (DT 2)
2..17	ASCII	Concentration range label

Command specific response codes for command 2469

Code	Type	Description
15	Error	Invalid Programmable Range Number
17	Error	Invalid Device Variable Index
19	Error	Device Variable Index Not Allowed For This Command

This command can only be used on Device Variables with Concentration Classification.

CombiLyz AFI4/AFI5 specific:

No check of invalid characters is performed on the label.

12.22 Command #2470 – Write Range Forced Status

The active range is often selectable using a combination on external pins.

This command sets up the active range to be forced to a certain range. Forced status is lost at restart.

Use NOT_USED to turn off forced mode.

This setting is volatile, and will be set to NOT_USED on reset!

Request data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)
1	uint-8	Forced range number (DT 2)

Response data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)
1	uint-8	Forced range number (DT 2)

Command specific response codes for command 2470

Code	Type	Description
15	Error	Invalid Programmable Range Number
17	Error	Invalid Device Variable Index
19	Error	Device Variable Index Not Allowed For This Command

CombiLyz AFI4/AFI5 specific:

This command forces all Device Variables to the specified range, no matter which Device Variable is targeted. Only on range is selectable at a time for all Device Variables.

13 Device-Specific Commands

The following device-specific commands are implemented:

- Command 130 – Read Device Variable State
- Command 131 – Write Device Variable State
- Command 132 – Read Device Variable Output Channel Setup
- Command 133 – Write Device Variable Output Channel Setup
- Command 134 – Reset Configuration to Factory Settings
- Command 135 – Trim Device Variable Zero
- Command 137 – Reset Device Variable Trim

13.1 Command #130 – Read Device Variable State

This command reads out the state whether the Device Variable is enabled or disabled.

Request data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)

Response data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)
1	uint-8	Device Variable State (DT 6)

Command specific response codes for command 130

Code	Type	Description
17	Error	Invalid Device Variable Index
19	Error	Device Variable Index Not Allowed For This Command

CombiLyz AFI4/AFI5 specific:

This command can only be used on Device Variables with Concentration Classification.

13.2 Command #131 – Write Device Variable State

This command writes the state whether the Device Variable is enabled or disabled.

Request data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)
1	uint-8	Device Variable State (DT 6)

Response data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)
1	uint-8	Device Variable State (DT 6)

Command specific response codes for command 131

Code	Type	Description
2	Error	Invalid Selection
17	Error	Invalid Device Variable Index
19	Error	Device Variable Index Not Allowed For This Command

CombiLyz AFI4/AFI5 specific:

This command is specifically used to enable and disable AFI4/AFI5's concentration mode.

This command can only be used on Device Variables with Concentration Classification.

Concentration enable is shared for Device Variable 2 and 4. If setting Concentration mode Enable for Device Variable 2, it will start outputting data (Outputs NOT_A_NUMBER when disabled), and Device Variable 4 will output Concentration instead of Conductivity.

This command cannot be used on Device Variable 4, since it is a multi-classification variable, and not Concentration classification.

13.3 Command #132 – Read Device Variable Output Channel Setup

This command reads the setup of the output for the specified device variable. This includes the current limits and the current to use in case of error, and which errors that may use this output as error channel.

Request data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)

Response data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)
1	uint-8	Output channel number (DT 8)
2..5	float	Lower channel output limit [mA]
6..9	float	Upper channel output limit [mA]
10	uint-8	Error output enable flags (DT 9)
11..14	float	Error current [mA]

Command specific response codes for command 132

Code	Type	Description
17	Error	Invalid Device Variable Index
19	Error	Device Variable Index Not Allowed For This Command

For device variables that cannot be assigned an output channel: Output Channel Number is set to NOT_USED, and floats are set to NOT_A_NUMBER. Error output flags are returned as zero.

13.4 Command #133 – Write Device Variable Output Channel Setup

This command sets up the Analog output channel source for the specified device variable and sets the limits which the output will not exceed in normal operation. It also sets up which errors are allowed to use this output channel, and what current an active error forces the output to.

Request data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)
1	uint-8	Output channel number (DT 8)
2..5	float	Lower channel output limit [mA]
6..9	float	Upper channel output limit [mA]
10	uint-8	Error output enable flags (DT 9)
11..14	float	Error current [mA]

Response data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)
1	uint-8	Output channel number (DT 8)
2..5	float	Lower channel output limit [mA]
6..9	float	Upper channel output limit [mA]
10	uint-8	Error output enable flags (DT 9)
11..14	float	Error current [mA]

Command specific response codes for command 133

Code	Type	Description
10	Error	Lower Limit Current Too Low
11	Error	Upper Limit Current Too High
15	Error	Invalid Error Current Value
17	Error	Invalid Device Variable Index
19	Error	Device Variable Index Not Allowed For This Command
29	Error	Invalid Span

CombiLyz AF14/AF15 specific:

The only allowed channel assignment is:

Device variable 1 – Temperature .. Channel 2

Device variable 4 – Combined Conductivity / Concentration .. Channel 1

No other device variables are allowed to have an assigned output channel.

Setting error current for either device variable 1 or 4, will set it for the other channel as well.

WARNING!: Error current is only checked for not overlapping the actual channel's limits – it may very well overlap another channel's limits.

13.5 Command #134 – Reset Configuration to Factory Settings

This command resets all configuration in the device to the state it was when shipped from the factory.

Request data frame

Byte	Format	Description
-	-	No request bytes

Response data frame

Byte	Format	Description
-	-	No response bytes

13.6 Command #135 – Trim Device Variable Zero

This command uses the provided parameter value and the actual measured device variable value and adjusts the internal zero compensation, so that the device variable output value matches the one sent with the command. It is possible to adjust the device variable zero with $\pm 5\%$ of the full range.

Request data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)
1	uint-8	Device Variable Unit (CT 2)
2..5	float	Externally measured value

Response data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)
1	uint-8	Device Variable Unit (CT 2)
2..5	float	Actual output Dev. Var. value

Command specific response codes for command 135

Code	Type	Description
3	Error	Passed Parameter Too Large
4	Error	Passed Parameter Too Small
17	Error	Invalid Device Variable Index
18	Error	Invalid Units Code
19	Error	Device Variable Index Not Allowed For This Command
29	Error	Invalid Span

CombiLyz AFI4/AFI5 specific:

This command is specifically designed for Device Variables 0 and 1.

13.7 Command #136 – Trim Device Variable Gain

This command is not implemented for CombiLyz AFI4/AFI5.

13.8 Command #137 – Reset Device Variable Trim

This command resets any user trim of the Device Variable to factory settings.

Request data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)

Response data frame

Byte	Format	Description
0	uint-8	Device Variable Code (DT 1)

Command specific response codes for command 137

Code	Type	Description
17	Error	Invalid Device Variable Index
19	Error	Device Variable Index Not Allowed For This Command

CombiLyz AFI4/AFI5 specific:

This command is specifically designed for Device Variables 0 and 1.

14 Conductivity Family Device Specific Tables

14.1 Cond. Family Device Spec. Table 1 – Conductivity Device Family Device Variable Status

Code	Description
0x01	Sensor Diagnostics Failure
0x02	Sensor Break (AFI4/AFI5specific)
0x04	Variable Not Generated (AFI4/AFI5 specific)
0x08	More Device Variable Status Available
0x30	Limit Status: 11 Constant (i.e. value cannot be changed by the process) 01 Low Limited (eg. A/D Converter has reached its lower limit) 10 High Limited (eg. A/D Converter has reached its upper limit) 00 Not Limited
0xC0	Process Data Quality Status: 11 Good 01 Poor accuracy (eg. value is beyond rated conductivity or hardware zoom, temperature out of range) 10 Manual / Fixed (eg. value is simulated or forced)

14.2 Conductivity Family Device Specific Table 2 – Conductivity Status 0

Code	Description
0x01	Conductivity Sensor Break
0x02	Process Temperature Sensor Break
0x04	Internal Temperature Sensor Break
0x08	Calibration Required
0x10	Conductivity Operating Range Exceeded
0x20	Process Temperature Operating Range Exceeded
0x40	Internal Temperature Operating Range Exceeded
0x80	Reserved

14.3 Conductivity Family Device Specific Table 3 – Conductivity Family Capabilities 0

Code	Description
0x01	Command 2307 – Read Conductivity Sensor Information
0x02	Command 2308 – Read Device Variable Range and Label Capabilities
0x04	Command 2309 – Read Conductivity Device Variables
0x08	Command 2435 – Write Conductivity Sensor Information
0x10	Reserved
0x20	Reserved
0x40	Reserved
0x80	Reserved

14.4 Conductivity Family Device Specific Table 4 - Conductivity Family Capabilities 1

Code	Description
0x01	Command 2336/2464 – Read/Write Range Hardware Range
0x02	Command 2337/2465 – Read/Write Range Output Zoom
0x04	Command 2338/2466 – Read/Write Range Temperature Compensation
0x08	Command 2339/2467 – Read/Write Range Conductivity Medium Label
0x10	Command 2340/2468 – Read/Write Range Concentration Table Points
0x20	Command 2341/2469 – Read/Write Range Concentration Medium Label
0x40	Command 2342/2470 – Read/Write Range Forced Status
0x80	Reserved

15 Device Specific Tables

15.1 Sensor Type Codes

Sensor Type codes

Code	Description
0	Contacting
1	Inductive
2	Electrode

15.2 Unit Codes

(subset of HART Common Table 2, Unit Codes)

Unit Codes

Code	Description
32	°C
33	°F
39	mA
57	%
66	mS/cm
67	µS/cm

15.3 Internal Unit Conversion

Conductivity Unit Conversion

$$\kappa[\mu\text{S}/\text{cm}] = \kappa[\text{mS}/\text{cm}] \times 1000$$

$$\kappa[\text{mS}/\text{cm}] = \kappa[\mu\text{S}/\text{cm}] / 1000$$

Temperature Unit Conversion

$$t[^\circ\text{C}] = (t[^\circ\text{F}] - 32) / 1.8$$

$$t[^\circ\text{F}] = t[^\circ\text{C}] \times 1.8 + 32$$

16 Performance

16.1 Sampling Rates

Typical sampling rates are shown in the following table.

Sensor	Sampling rate
Conductivity	Approx. 3 times per second
Temperature	Approx. 1 time per second

Note: both temperature calculations use an equally-weighted running mean of the last 3 input values.

16.2 Power-Up

On power up, the transmitter runs through a startup initialization procedure, which takes approximately 5 seconds. During this period, the device will not be able to respond to HART commands, and the analog output is set at 3.5mA.

The first stable measurements are ready in less than 15 seconds, allowing valid Device Variable readouts. Fixed-current mode is cancelled upon startup / reset.

16.3 Reset

Command 42 - Perform Device Reset causes the device to reset its microprocessor. The resulting restart is identical to the normal power up sequence.

16.4 Self-Test

The CombiLyz AFI4/AFI5 does not support command 41 – Self Test. Self-testing is performed periodically during normal operation.

16.5 Command Response Times

HART command response time depends on the command number issued and the internal state of the device. If write commands result in writing in non-volatile memory, the response is sent upon completion of the write, causing a small delay of up to 140ms.

Generalization	Response times
Minimum	5ms
Typical	15ms
Maximum	140ms

16.6 Busy and Delayed-Response

Delayed-response is not used.

16.7 Long Messages

The largest data field used is in the response to Command 2468, where 39 bytes (not including the two status bytes) are returned. Total length of data field is then 41 bytes.

16.8 Non-Volatile Memory

Built-in flash memory of the micro controller is used to hold user configuration. New data is written to this memory immediately on execution of a write command, before the response is sent. If the new data to be written is identical to the data already stored, the write is not performed.

16.9 Modes

Fixed current mode is implemented, using Command 40. This mode is cleared at startup or reset.

16.10 Burst Mode

This Field Device does not support Burst Mode.

16.11 Write Protection

This Field Device does not support Write Protect.

16.12 Catch Device Variable

This Field Device does not support Catch Device Variable.

16.13 Damping

Damping is implemented for the PV and the other device variables depending on the PV value. This includes the conductivity/concentration output channel current, and if a concentration value is calculated (which always is done with respect to the PV value).

ANNEX A. CAPABILITY CHECKLIST

Manufacturer, model and revision	Baumer CombiLyz AF14/AF15 HART
Device type	Conductivity Current Output Transmitter
HART revision	7.4
Device Description available	Yes
Number and type of sensors	One internal conductivity sensor One internal temperature sensor
Number and type of actuators	None
Number and type of host side signals	One 4 - 20mA analog output w/ HART One 4 - 20mA analog output
Number of Device Variables	8
Number of Dynamic Variables	4
Mapable Dynamic Variables?	No
Number of common-practice commands	14
Number of device-specific commands	7
Bytes of additional device status	14
Alternative operating modes?	No
Burst mode?	No
Write-protection?	No

ANNEX B. DEFAULT CONFIGURATION

Parameter	Default value
Poll Address	0
Loop Current Signaling Mode	Enabled
Conductivity / Concentration Mode	Conductivity Mode
Conductivity Current at 4 mA	Range 1..4: 0 μ S/cm
Conductivity Current at 20 mA	Range 1: 200 mS/cm Range 2: 20 mS/cm Range 3: 2 mS/cm Range 4: 500 μ S/cm
Temperature Current at 4 mA	-20 $^{\circ}$ C
Temperature Current at 20 mA	150 $^{\circ}$ C
Temperature compensation	Range 1..4: 2 % / K
Error Channel	Both
Error Current	3.5 mA
Damping	0 sec

NOTE: By default, all strings are cleared (all space characters)

ANNEX C. REVISION HISTORY

A1. 2015-09-02 First Revision 1.0

Document created.